



# सीएसआईआर-केन्द्रीय इलेक्ट्रॉनिकी अभियांत्रिकी अनुसंधान संस्थ

CSIR-CENTRAL ELECTRONICS ENGINEERING RESEARCH INSTITUTE

(विज्ञान तथा प्रौद्योगिकी मंत्रालय / MINISTRY OF SCIENCE & TECHNOLOGY, भारत सरकार / GOVT. OF INDIA)

पिलानी, राजस्थान (भारत)/Pilani, Rajasthan - 333031 (INDIA)



## Corrigendum

Tender ID: 2018\_CSIR\_17337\_1

Tender Reference No: 468/GAP3129/IIH/Jaipur/8Pur/2018

Tender Title: Power Quality Analyzer 6 Channel and External Current Probes

Tender Type: Global Tender Inquiry

This corrigendum is in reference to the above mentioned details and the following information/ amendments shall hold true with reference to the Detailed Technical Specifications

- **6 Channel Power Quality Analyzer**

<i>Technical specifications of power analyser</i>		
Sl. No.	Detailed Specification	
02.	Voltage Range	10 V to 1000V

- **Compatible current Measuring Probes with 6 Channel Power Quality Analyzer**

<i>Technical specifications of Compatible Current Measuring Probes with 6 Channel Power Quality Analyzer</i>		
Sl. No.	Detailed Specification	
03.	Operating Temperature	10 to 40 degree Celcius

Stores & Purchase Officer

*It is to be kindly noted that this Corrigendum supercede the previously published Corrigendum*