

CORRIGENDUM

In continuation to this office GLOBAL TENDER advertisement on CPP Portal ID NO. 2021_CSIR_98011_1 dated 11.12.2021 and Tender No. 326/MLP-0121/VEDDG/10-Pur/2021/T-95 dated 11.12.2021 on CEERI Website for the purchase of 4GHz 4Channel Digital Oscilloscope. Please read the amended specifications at Sl. No. 9, 10, 12 and 17 mentioned below.

Sl. No.	Specification mentioned in tender document	Amended Specifications
9.	Number of bits	8 bit ADC or better
10.	Vertical sensitivity (per division)	1 mV/div to 1V/div @50Ω and @1MΩ
12.	Timebase accuracy	±100 ppb or better
17.	Waveform Update Rate (Waveform/sec)	≥1 000 0d0 waveforms/s

All other specifications will as per the original tender document.

Stores & Purchase Officer
CSIR-CEERI, Pilani