

CORRIGENDUM

In continuation to this office advertisement through CPP Portal as EXPRESSION OF INTEREST (EOI) ID NO. 2023_CSIR_162359_1 and Expression of Interest No. SPO/EOI/Jaipur Centre/10-Pur/2023/03/02 dated 26.07.2023 for the Technical Proposal, discussion and presentation of High Voltage DC Tester In this regard this is to inform you that please read the specifications of High Voltage DC Tester mentioned below:

Sr. No.	Parameters	Value
1.	Input Power Supply	Single Phase Mains 230V, 50Hz & Internal Rechargeable Battery
2.	Output Voltage	0 – 25kV DC or higher
3.	Output Rated Current	1.5mA or higher (at Maximum Voltage)
4.	Battery Operation Time at Full Load	20 min or more
5.	Voltage Measurement Range	0 25kV or higher
6.	Current Measurement Range	0 – 1.5 mA or higher
7.	Accessories	Protective Bag
		Mains Lead
		Ground Cable
		HV Connecting Leads
		Start Keys

Stores & Purchase Officer
CSIR-CEERI, Pilani